
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000773
12:34 09/11/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:30
PRESSURIZING 3:30
STERILIZING 4:34
136.4 C 224kPa 4:34
136.3 C 224kPa 5:04
136.4 C 224kPa 5:34
136.4 C 224kPa 6:04
136.1 C 224kPa 6:34
136.3 C 222kPa 7:04
136.4 C 220kPa 7:34
136.4 C 222kPa 8:04
136.4 C 222kPa 8:04
VENTING 8:04
AIR DRYING 9:00
CYCLE COMPLETE 9:06

Digital Signature #
93F0597118FBFDC5
