
Midmark M11 - v1.0.0
Total Cycles: 4304

Name: RITTER 1
Sterilizer ID: V2178198
02/15/2024 09:40 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.2 F 0.0
2:00 83.2 F 0.1
4:00 85.7 F 0.8
6:00 112.1 F 2.8
8:00 142.5 F 5.1
10:00 167.7 F 7.6
12:00 182.3 F 10.5
14:00 197.0 F 10.8
16:00 212.1 F 4.1
18:00 219.2 F 6.9
20:00 229.3 F 10.3
22:00 233.2 F 9.8
24:00 236.1 F 9.4
26:00 241.7 F 11.1
28:00 243.7 F 11.9
30:00 254.2 F 17.2
32:00 263.7 F 22.9
34:00 272.7 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.0
1:00 271.9 F 28.6
2:00 272.1 F 28.3
3:00 271.2 F 28.2
4:00 272.2 F 28.7

Min 271.0 F 27.9
Max 272.6 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 179.8 F 0.0
5:00 236.0 F 0.0
10:00 222.9 F 0.0
15:00 201.8 F 0.0
20:00 178.1 F 0.0
25:00 175.6 F 0.0
30:00 173.2 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 34:13
STERILIZING: 4:00
VENTING: 3:38
DRYING: 30:00
TOTAL CYCLE: 01:12:42
CYCLE COMPLETE

02/15/2024 10:52 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____