

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4870

Name: RITTER 1  
Sterilizer ID: V2178198  
09/05/2024 12:42 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 99.9 F 0.0  
2:00 102.8 F 0.0  
4:00 139.5 F 0.4  
6:00 160.7 F 0.6  
8:00 179.1 F 2.8  
10:00 196.4 F 5.9  
12:00 218.2 F 9.5  
14:00 222.6 F 12.4  
16:00 231.7 F 8.3  
18:00 242.9 F 13.1  
20:00 239.3 F 10.1  
22:00 236.4 F 8.4  
24:00 248.5 F 14.1  
26:00 259.7 F 20.3  
28:00 269.8 F 26.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 28.3  
1:00 272.6 F 28.6  
2:00 272.7 F 29.0  
3:00 271.7 F 28.3  
4:00 272.7 F 28.7

Min 271.7 F 28.2  
Max 272.9 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 187.6 F 0.0  
5:00 208.6 F 0.0  
10:00 191.7 F 0.0  
15:00 203.7 F 0.0  
20:00 212.4 F 0.0  
25:00 220.1 F 0.0  
30:00 222.5 F 0.0  
Drying Complete

FILLING: 0:46  
HEATING: 29:41  
STERILIZING: 4:00  
VENTING: 3:41  
DRYING: 30:00  
TOTAL CYCLE: 01:08:08  
CYCLE COMPLETE

09/05/2024 01:50 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_