
Midmark M11 - v1.0.5
Total Cycles: 2932

Name: RITTER2
Sterilizer ID: V2313228
09/26/2023 02:41 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 169.6 F 0.0
2:00 175.6 F 3.3
4:00 194.1 F 7.8
6:00 212.6 F 12.5
8:00 225.0 F 7.5
10:00 239.3 F 12.8
12:00 239.2 F 10.6
14:00 236.8 F 9.3
16:00 251.1 F 16.0
18:00 262.5 F 22.6
20:00 272.1 F 29.6

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.9
1:00 272.8 F 29.9
2:00 272.3 F 29.3
3:00 271.1 F 28.8
4:00 272.0 F 29.6

Min 271.0 F 28.8
Max 273.3 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.9 F 0.0
5:00 218.7 F 0.0
10:00 217.3 F 0.0
15:00 187.2 F 0.0
20:00 179.0 F 0.0
25:00 176.1 F 0.0
30:00 175.8 F 0.0
Drying Complete

FILLING: 1:49
HEATING: 20:31
STERILIZING: 4:00
VENTING: 3:41
DRYING: 30:00
TOTAL CYCLE: 01:00:01
CYCLE COMPLETE

09/26/2023 03:40 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____