
Midmark M11 - v1.0.0
Total Cycles: 1227

Name: RITTER 1
Sterilizer ID: V2178198
10/04/2021 09:15 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 139.7 F 0.0
2:00 139.8 F 0.9
4:00 155.3 F 3.0
6:00 185.5 F 6.7
8:00 201.6 F 11.1
10:00 213.1 F 4.4
12:00 227.5 F 9.0
14:00 237.8 F 12.4
16:00 238.7 F 10.8
18:00 236.0 F 8.8
20:00 250.0 F 15.4
22:00 261.4 F 21.8
24:00 271.9 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.5
1:00 273.0 F 29.2
2:00 271.8 F 28.4
3:00 272.5 F 29.2
4:00 272.6 F 28.9

Min 271.6 F 28.4
Max 273.0 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.7 F 0.0
5:00 219.9 F 0.0
10:00 206.6 F 0.0
15:00 197.1 F 0.0
20:00 175.7 F 0.0
25:00 196.5 F 0.0
Dry Cycle Aborted

FILLING: 1:04
HEATING: 24:35
STERILIZING: 4:00
VENTING: 3:21
DRYING: 29:44
TOTAL CYCLE: 01:02:44

10/04/2021 10:17 AM
ACCEPTED: X
REJECTED:
Operator: U01