

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 002568  
12:57 07/03/2025  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:02  
PRESSURIZING 2:10  
STERILIZING 3:17  
136.5 C 224kPa 3:17  
136.4 C 222kPa 3:47  
136.3 C 224kPa 4:17  
136.4 C 223kPa 4:47  
136.0 C 221kPa 5:17  
136.1 C 222kPa 5:47  
136.4 C 223kPa 6:17  
136.3 C 224kPa 6:47  
136.3 C 224kPa 6:47  
VENTING 6:48  
AIR DRYING 7:40  
CYCLE COMPLETE 37:40  
-----

Digital Signature #  
861D34EAB3C547C5  
-----