
Midmark M11 - v1.0.5
Total Cycles: 3631

Name: RITTER2
Sterilizer ID: V2313228
05/21/2024 11:42 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 140.9 F 0.0
2:00 139.1 F 1.6
4:00 146.4 F 4.3
6:00 165.2 F 7.8
8:00 187.1 F 12.0
10:00 206.8 F 5.1
12:00 223.3 F 9.2
14:00 233.8 F 12.9
16:00 234.8 F 10.6
18:00 235.2 F 8.4
20:00 246.6 F 14.4
22:00 257.1 F 20.2
24:00 266.9 F 26.3
26:00 272.5 F 30.3

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 30.0
1:00 271.7 F 30.0
2:00 271.5 F 30.0
3:00 271.6 F 29.9
4:00 271.8 F 29.9

Min 271.4 F 29.9
Max 272.9 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.3 F 0.0
5:00 233.9 F 0.0
10:00 223.4 F 0.0
15:00 206.1 F 0.0
20:00 201.8 F 0.0
25:00 196.5 F 0.0
30:00 191.3 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 26:23
STERILIZING: 4:00
VENTING: 3:37
DRYING: 30:00
TOTAL CYCLE: 01:04:55
CYCLE COMPLETE

05/21/2024 12:46 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____