
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001877
9:58 22/07/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:51
PRESSURIZING 2:57
STERILIZING 4:07
136.0 C 222kPa 4:07
136.3 C 224kPa 4:37
136.4 C 224kPa 5:07
136.3 C 224kPa 5:37
136.4 C 225kPa 6:07
136.2 C 223kPa 6:37
136.2 C 223kPa 7:07
136.3 C 225kPa 7:37
136.3 C 225kPa 7:37
VENTING 7:38
AIR DRYING 8:28
CYCLE COMPLETE 9:00

Digital Signature #
6E603D6829B5EDDF
