
Midmark M11 - v1.0.0
Total Cycles: 2143

Name: RITTER 1
Sterilizer ID: V2178198
07/18/2022 10:23 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 189.0 F 0.0
2:00 186.6 F 2.3
4:00 185.0 F 6.4
6:00 218.8 F 11.1
8:00 224.0 F 6.1
10:00 234.8 F 11.3
12:00 233.7 F 8.0
14:00 241.1 F 11.3
16:00 245.0 F 12.8
18:00 256.9 F 19.4
20:00 268.4 F 26.4
22:00 271.8 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.0
1:00 272.8 F 29.4
2:00 272.0 F 29.3
3:00 272.7 F 29.4
4:00 272.0 F 29.2

Min 271.8 F 28.9
Max 272.8 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.3 F 0.0
5:00 231.2 F 0.0
10:00 243.9 F 0.0
15:00 236.9 F 0.0
20:00 224.4 F 0.0
25:00 219.7 F 0.0
30:00 224.0 F 0.0
Drying Complete

FILLING: 1:09
HEATING: 22:01
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:00:22
CYCLE COMPLETE

07/18/2022 11:23 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____