
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000049
13:38 28/02/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:47
PRESSURIZING 2:58
STERILIZING 4:10
136.0 C 221kPa 4:10
136.1 C 223kPa 4:40
136.1 C 222kPa 5:10
136.1 C 223kPa 5:40
136.3 C 224kPa 6:10
136.1 C 223kPa 6:40
136.4 C 226kPa 7:10
136.4 C 225kPa 7:40
136.4 C 225kPa 7:40
VENTING 7:40
CYCLE ABORTED 8:01
STOP BUTTON PRESSED
NOT STERILE
END TIME 8:01

Digital Signature #
128D4312DC051D4D
