
Midmark M11 - v1.0.0
Total Cycles: 4213

Name: RITTER 1
Sterilizer ID: V2178198
01/19/2024 09:06 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.2 F 0.0
2:00 69.3 F 0.2
4:00 86.1 F 1.1
6:00 106.7 F 2.9
8:00 139.2 F 5.3
10:00 163.4 F 8.1
12:00 191.5 F 11.5
14:00 210.7 F 2.7
16:00 221.4 F 5.4
18:00 227.0 F 8.6
20:00 236.7 F 12.1
22:00 230.9 F 7.2
24:00 241.2 F 12.0
26:00 237.6 F 9.3
28:00 249.8 F 15.1
30:00 260.0 F 20.8
32:00 269.6 F 26.9

STERILIZING
mm:ss Degrees PSI
0:00 272.3 F 28.6
1:00 272.8 F 29.2
2:00 271.7 F 28.7
3:00 272.4 F 28.8
4:00 272.3 F 29.2

Min 271.4 F 28.5
Max 273.0 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.2 F 0.0
5:00 231.1 F 0.0
10:00 220.5 F 0.0
15:00 222.1 F 0.0
20:00 226.8 F 0.0
25:00 228.2 F 0.0
30:00 227.0 F 0.0

Drying Complete

FILLING: 0:52
HEATING: 33:37
STERILIZING: 4:00
VENTING: 3:37
DRYING: 30:00
TOTAL CYCLE: 01:12:07
CYCLE COMPLETE

01/19/2024 10:18 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____