
Midmark M11 - v1.0.5
Total Cycles: 3071

Name: RITTER2
Sterilizer ID: V2313228
11/01/2023 01:03 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 129.6 F 0.0
2:00 130.9 F 0.9
4:00 150.2 F 3.7
6:00 176.9 F 7.7
8:00 195.8 F 12.4
10:00 218.6 F 5.9
12:00 226.4 F 10.6
14:00 229.7 F 7.4
16:00 242.9 F 12.9
18:00 242.1 F 11.6
20:00 254.7 F 18.2
22:00 266.1 F 25.1

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.8
1:00 272.2 F 29.6
2:00 271.7 F 28.8
3:00 272.0 F 29.6
4:00 272.1 F 29.1

Min 271.1 F 28.7
Max 272.4 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.5 F 0.0
5:00 235.8 F 0.0
10:00 192.9 F 0.0
15:00 191.4 F 0.0
20:00 187.6 F 0.0
25:00 184.6 F 0.0
30:00 182.4 F 0.0
Drying Complete

FILLING: 1:02
HEATING: 23:31
STERILIZING: 4:00
VENTING: 3:25
DRYING: 30:00
TOTAL CYCLE: 01:01:58
CYCLE COMPLETE

11/01/2023 02:04 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____