
Midmark M11 - v1.0.5
Total Cycles: 2297

Name: RITTER2
Sterilizer ID: V2313228
04/25/2023 08:50 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 153.4 F 0.0
2:00 151.9 F 1.7
4:00 155.9 F 4.5
6:00 183.7 F 8.7
8:00 210.1 F 9.3
10:00 221.4 F 8.2
12:00 233.7 F 13.5
14:00 237.2 F 10.5
16:00 235.7 F 8.8
18:00 249.9 F 15.7
20:00 261.9 F 22.6
22:00 272.9 F 29.6

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.0
1:00 272.5 F 29.6
2:00 271.3 F 28.9
3:00 272.2 F 29.8
4:00 272.2 F 29.4

Min 271.1 F 28.9
Max 272.6 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.4 F 0.0
5:00 202.6 F 0.0
10:00 210.4 F 0.0
15:00 239.5 F 0.0
20:00 229.0 F 0.0
25:00 230.4 F 0.0
30:00 230.6 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 22:23
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:00:31
CYCLE COMPLETE

04/25/2023 09:50 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____