
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 004383
13:41 19/09/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:08
PRESSURIZING 2:11
STERILIZING 3:10
136.4 C 224kPa 3:10
136.0 C 222kPa 3:40
136.1 C 222kPa 4:10
136.2 C 224kPa 4:40
136.1 C 222kPa 5:10
136.4 C 225kPa 5:40
136.4 C 224kPa 6:10
136.1 C 222kPa 6:40
136.1 C 222kPa 6:40
VENTING 6:41
AIR DRYING 8:43
CYCLE COMPLETE 8:51

Digital Signature #
F99F7302FC634713
