
Midmark M11 - v1.0.0
Total Cycles: 1439

Name: RITTER 1
Sterilizer ID: V2178198
12/09/2021 09:04 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 168.4 F 0.0
2:00 160.6 F 1.3
4:00 175.3 F 4.4
6:00 195.7 F 8.6
8:00 217.7 F 8.6
10:00 230.4 F 8.5
12:00 238.9 F 12.1
14:00 240.7 F 10.9
16:00 237.2 F 9.0
18:00 252.1 F 16.0
20:00 263.9 F 23.0
22:00 271.3 F 27.8

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.0
1:00 272.7 F 29.0
2:00 272.2 F 28.3
3:00 271.2 F 28.0
4:00 272.6 F 29.0

Min 271.1 F 27.8
Max 273.0 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.5 F 0.0
5:00 228.4 F 0.0
10:00 219.7 F 0.0
15:00 220.6 F 0.0
20:00 219.7 F 0.0
25:00 228.0 F 0.0
30:00 237.0 F 0.0
Drying Complete

FILLING: 1:26
HEATING: 22:05
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 01:00:38
CYCLE COMPLETE

12/09/2021 10:15 AM
ACCEPTED: X
REJECTED:
Operator: U02