

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2494

Name: RITTER 1  
Sterilizer ID: V2178198  
10/26/2022 02:01 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 194.6 F 0.0  
2:00 189.7 F 4.5  
4:00 208.3 F 9.4  
6:00 219.0 F 6.8  
8:00 232.3 F 9.6  
10:00 235.8 F 8.5  
12:00 243.7 F 13.1  
14:00 243.1 F 11.9  
16:00 255.7 F 18.7  
18:00 267.4 F 25.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.7  
1:00 272.4 F 29.3  
2:00 271.5 F 28.7  
3:00 272.4 F 29.4  
4:00 271.7 F 28.7  
  
Min 271.3 F 28.6  
Max 272.5 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.6 F 0.0  
5:00 237.5 F 0.0  
10:00 229.1 F 0.0  
15:00 228.9 F 0.0  
20:00 230.4 F 0.0  
25:00 225.5 F 0.0  
30:00 225.7 F 0.0  
Drying Complete

FILLING: 1:29  
HEATING: 19:20  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 00:58:03  
CYCLE COMPLETE

10/26/2022 02:59 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_