
Midmark M11 - v1.0.5
Total Cycles: 4204

Name: RITTER2
Sterilizer ID: V2313228
12/11/2024 01:21 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 97.9 F 0.0
2:00 98.4 F 0.0
4:00 103.8 F 0.9
6:00 136.6 F 3.6
8:00 161.3 F 6.6
10:00 186.4 F 9.9
12:00 207.7 F 9.1
14:00 222.2 F 7.7
16:00 234.2 F 12.2
18:00 233.6 F 9.1
20:00 239.0 F 10.9
22:00 243.5 F 12.9
24:00 255.2 F 18.8
26:00 265.5 F 25.1
28:00 271.2 F 29.7

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.9
1:00 272.7 F 30.3
2:00 271.7 F 30.2
3:00 272.2 F 30.2
4:00 271.7 F 30.3

Min 271.4 F 29.8
Max 272.9 F 30.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.0 F 0.0
5:00 227.3 F 0.0
10:00 203.6 F 0.0
15:00 196.8 F 0.0
20:00 197.9 F 0.0
25:00 189.9 F 0.0
30:00 189.3 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 28:38
STERILIZING: 4:00
VENTING: 3:48
DRYING: 30:00
TOTAL CYCLE: 01:07:25
CYCLE COMPLETE

12/11/2024 02:28 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____