
Midmark M11 - v1.0.5
Total Cycles: 787

Name: RITTER2
Sterilizer ID: V2313228
02/10/2022 06:58 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 166.2 F 0.0
2:00 169.6 F 3.4
4:00 192.7 F 8.2
6:00 213.6 F 11.2
8:00 226.8 F 9.0
10:00 234.7 F 9.9
12:00 242.5 F 13.1
14:00 243.1 F 12.4
16:00 256.1 F 19.3
18:00 267.6 F 26.4

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.1
1:00 272.7 F 29.8
2:00 271.5 F 29.0
3:00 272.7 F 30.0
4:00 272.2 F 29.4

Min 271.3 F 29.0
Max 272.9 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 177.8 F 0.0
5:00 235.2 F 0.0
10:00 195.3 F 0.0
15:00 189.3 F 0.0
20:00 186.1 F 0.0
25:00 185.4 F 0.0
30:00 184.3 F 0.0
Drying Complete

FILLING: 2:05
HEATING: 19:24
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:58:40
CYCLE COMPLETE

02/10/2022 07:56 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____