

-----  
Midmark M11 - v1.0.0  
Total Cycles: 684

Name: MIDMARK  
Sterilizer ID: V2178198  
02/22/2021 08:47 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 197.0 F 0.0  
2:00 194.2 F 1.9  
4:00 206.9 F 5.2  
6:00 208.7 F 10.9  
8:00 223.0 F 4.8  
10:00 233.3 F 9.8  
12:00 238.6 F 9.4  
14:00 244.0 F 12.5  
16:00 241.8 F 11.0  
18:00 254.3 F 17.4  
20:00 265.2 F 23.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.1  
1:00 272.6 F 28.7  
2:00 271.2 F 28.1  
3:00 272.7 F 28.8  
4:00 271.5 F 28.0  
  
Min 271.2 F 27.9  
Max 272.8 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 205.1 F 0.0  
5:00 214.9 F 0.0  
10:00 235.6 F 0.0  
15:00 233.6 F 0.0  
20:00 226.6 F 0.0  
25:00 226.0 F 0.0  
30:00 172.1 F 0.0  
Drying Complete

FILLING: 1:36  
HEATING: 21:53  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 01:00:40  
CYCLE COMPLETE

02/22/2021 09:47 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01