
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000806
12:19 04/10/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:41
PRESSURIZING 2:46
STERILIZING 4:13
136.4 C 231kPa 4:13
136.4 C 233kPa 4:43
136.4 C 233kPa 5:13
136.4 C 233kPa 5:43
136.3 C 233kPa 6:13
136.4 C 233kPa 6:43
136.4 C 233kPa 7:13
136.3 C 233kPa 7:43
136.3 C 233kPa 7:43
VENTING 7:43
AIR DRYING 8:38
CYCLE COMPLETE 8:44

Digital Signature #
0F5FBE63D1013684
