
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 004797
10:58 05/12/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:00
PRESSURIZING 3:04
STERILIZING 4:17
136.4 C 220kPa 4:17
136.3 C 224kPa 4:47
136.4 C 223kPa 5:17
136.2 C 223kPa 5:47
136.4 C 224kPa 6:17
136.2 C 218kPa 6:47
136.3 C 224kPa 7:17
136.3 C 222kPa 7:47
136.3 C 222kPa 7:47
VENTING 7:47
AIR DRYING 8:45
CYCLE COMPLETE 15:06

Digital Signature #
1443152C414304ED
