
Midmark M11 - v1.0.5
Total Cycles: 1573

Name: RITTER2
Sterilizer ID: V2313228
10/06/2022 09:45 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 153.8 F 0.0
2:00 153.3 F 0.0
4:00 164.1 F 0.4
6:00 191.9 F 4.1
8:00 209.6 F 8.9
10:00 228.1 F 9.4
12:00 233.9 F 10.6
14:00 234.8 F 8.3
16:00 241.3 F 10.6
18:00 247.0 F 14.3
20:00 259.3 F 21.1
22:00 270.4 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.1
1:00 272.8 F 29.7
2:00 271.8 F 29.2
3:00 273.0 F 29.7
4:00 271.9 F 29.0

Min 271.8 F 29.0
Max 273.0 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.3 F 0.0
5:00 200.9 F 0.0
10:00 212.1 F 0.0
15:00 214.4 F 0.0
20:00 214.6 F 0.0
25:00 214.2 F 0.0
30:00 213.7 F 0.0
Drying Complete

FILLING: 1:09
HEATING: 22:50
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:01:10
CYCLE COMPLETE

10/06/2022 10:46 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____