

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2661

Name: RITTER 1  
Sterilizer ID: V2178198  
12/08/2022 01:23 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 206.0 F 0.0  
2:00 196.3 F 4.1  
4:00 210.3 F 9.1  
6:00 222.8 F 8.4  
8:00 231.7 F 10.1  
10:00 235.7 F 7.9  
12:00 245.4 F 12.9  
14:00 244.8 F 12.5  
16:00 258.0 F 19.5  
18:00 269.5 F 26.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.3  
1:00 272.6 F 29.0  
2:00 271.7 F 28.2  
3:00 272.5 F 29.1  
4:00 272.2 F 28.7  
  
Min 271.5 F 28.1  
Max 272.7 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.5 F 0.0  
5:00 232.2 F 0.0  
10:00 232.0 F 0.0  
15:00 233.9 F 0.0  
20:00 220.8 F 0.0  
25:00 218.2 F 0.0  
30:00 222.3 F 0.0  
Drying Complete

FILLING: 1:40  
HEATING: 19:02  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 00:57:51  
CYCLE COMPLETE

12/08/2022 02:20 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_