
Midmark M11 - v1.0.5
Total Cycles: 511

Name: RITTER2
Sterilizer ID: V2313228
11/02/2021 06:32 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 72.9 F 0.0
2:00 73.4 F 0.0
4:00 76.8 F 0.0
6:00 93.2 F 0.1
8:00 133.0 F 0.3
10:00 180.8 F 0.4
12:00 200.5 F 3.5
14:00 213.1 F 7.2
16:00 225.7 F 11.0
18:00 229.3 F 7.5
20:00 241.6 F 13.0
22:00 241.1 F 11.7
24:00 240.9 F 11.0
26:00 254.7 F 18.0
28:00 266.3 F 25.1

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.7
1:00 272.1 F 29.1
2:00 271.7 F 29.1
3:00 271.8 F 28.9
4:00 271.7 F 29.2

Min 271.2 F 28.5
Max 272.7 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.8 F 0.0
5:00 223.1 F 0.0
10:00 232.3 F 0.0
15:00 223.9 F 0.0
20:00 219.8 F 0.0
25:00 221.6 F 0.0
30:00 221.6 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 29:34
STERILIZING: 4:00
VENTING: 2:55
DRYING: 30:00
TOTAL CYCLE: 01:07:25
CYCLE COMPLETE

11/02/2021 07:39 AM
ACCEPTED: X
REJECTED:
Operator: U01