
Midmark M11 - v1.0.0
Total Cycles: 3715

Name: RITTER 1
Sterilizer ID: V2178198
09/05/2023 09:23 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 190.5 F 0.0
2:00 191.2 F 0.4
4:00 186.8 F 3.5
6:00 205.9 F 8.6
8:00 223.7 F 9.7
10:00 230.8 F 9.0
12:00 238.4 F 10.3
14:00 243.1 F 12.3
16:00 241.4 F 11.0
18:00 254.7 F 17.7
20:00 266.2 F 24.7

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.3
1:00 272.5 F 28.8
2:00 271.8 F 28.5
3:00 272.4 F 28.7
4:00 271.7 F 28.4

Min 271.6 F 28.1
Max 272.8 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.0 F 0.0
5:00 221.1 F 0.0
10:00 226.2 F 0.0
15:00 226.6 F 0.0
20:00 226.9 F 0.0
25:00 227.4 F 0.0
30:00 226.8 F 0.0
Drying Complete

FILLING: 1:17
HEATING: 21:35
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 01:00:05
CYCLE COMPLETE

09/05/2023 10:23 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____