
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 006220
13:50 31/08/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:38
PRESSURIZING 2:40
STERILIZING 3:44
136.4 C 223kPa 3:44
136.2 C 223kPa 4:15
136.2 C 220kPa 4:45
136.4 C 225kPa 5:15
136.3 C 222kPa 5:45
136.4 C 224kPa 6:15
136.4 C 223kPa 6:45
136.4 C 224kPa 7:15
136.4 C 224kPa 7:15
VENTING 7:15
AIR DRYING 8:04
CYCLE COMPLETE 31:44

Digital Signature #
7C62E2C2B66F8B64
