

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4199

Name: RITTER 1  
Sterilizer ID: V2178198  
01/12/2024 12:44 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 153.2 F 0.0  
2:00 154.4 F 2.6  
4:00 168.4 F 6.0  
6:00 194.8 F 9.8  
8:00 219.1 F 5.2  
10:00 221.1 F 7.9  
12:00 233.0 F 12.7  
14:00 234.9 F 9.3  
16:00 239.1 F 10.0  
18:00 246.3 F 13.4  
20:00 258.2 F 19.8  
22:00 268.9 F 26.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.3  
1:00 272.6 F 28.9  
2:00 271.1 F 28.2  
3:00 272.4 F 29.1  
4:00 271.8 F 28.6

Min 271.0 F 28.2  
Max 272.7 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 204.1 F 0.0  
5:00 229.0 F 0.0  
10:00 237.6 F 0.0  
15:00 236.0 F 0.0  
20:00 225.9 F 0.0  
25:00 226.7 F 0.0  
30:00 228.4 F 0.0  
Drying Complete

FILLING: 1:09  
HEATING: 23:09  
STERILIZING: 4:00  
VENTING: 3:37  
DRYING: 30:00  
TOTAL CYCLE: 01:01:55  
CYCLE COMPLETE

01/12/2024 01:45 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_