
Midmark M11 - v1.0.5
Total Cycles: 2109

Name: RITTER2
Sterilizer ID: V2313228
03/01/2023 12:55 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 202.1 F 0.0
2:00 201.3 F 3.9
4:00 205.3 F 8.6
6:00 215.0 F 9.5
8:00 229.8 F 9.2
10:00 234.9 F 9.3
12:00 242.5 F 12.7
14:00 242.1 F 11.7
16:00 255.2 F 18.6
18:00 266.7 F 25.7

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.1
1:00 272.4 F 29.8
2:00 271.6 F 29.0
3:00 271.9 F 29.8
4:00 272.4 F 29.6

Min 271.2 F 28.9
Max 272.6 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.2 F 0.0
5:00 223.2 F 0.0
10:00 231.1 F 0.0
15:00 232.2 F 0.0
20:00 227.6 F 0.0
25:00 228.4 F 0.0
30:00 229.0 F 0.0
Drying Complete

FILLING: 1:41
HEATING: 19:32
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:58:27
CYCLE COMPLETE

03/01/2023 01:53 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____