
Midmark M11 - v1.0.5
Total Cycles: 3545

Name: RITTER2
Sterilizer ID: V2313228
04/22/2024 10:55 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 154.3 F 0.0
2:00 147.1 F 1.6
4:00 154.7 F 4.2
6:00 177.9 F 7.8
8:00 199.6 F 11.8
10:00 218.9 F 5.0
12:00 229.4 F 9.3
14:00 236.6 F 11.1
16:00 239.3 F 11.0
18:00 235.9 F 9.1
20:00 249.6 F 15.3
22:00 260.5 F 21.5
24:00 270.6 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 28.8
1:00 272.7 F 29.4
2:00 271.4 F 29.1
3:00 272.0 F 28.9
4:00 272.3 F 29.5

Min 271.1 F 28.7
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.0 F 0.0
5:00 239.4 F 0.0
10:00 226.1 F 0.0
15:00 202.6 F 0.0
20:00 220.8 F 0.0
25:00 226.1 F 0.0
30:00 226.4 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 24:39
STERILIZING: 4:00
VENTING: 3:26
DRYING: 30:00
TOTAL CYCLE: 01:03:02
CYCLE COMPLETE

04/22/2024 11:57 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____