
Midmark M11 - v1.0.5
Total Cycles: 2894

Name: RITTER2
Sterilizer ID: V2313228
09/15/2023 05:15 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 165.6 F 0.0
2:00 163.5 F 3.1
4:00 180.9 F 8.0
6:00 204.9 F 12.9
8:00 222.8 F 8.5
10:00 233.4 F 10.5
12:00 239.7 F 12.1
14:00 240.1 F 11.0
16:00 253.5 F 17.8
18:00 265.1 F 24.8

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.1
1:00 272.7 F 29.8
2:00 271.3 F 29.1
3:00 272.7 F 29.8
4:00 271.5 F 29.1

Min 271.3 F 29.0
Max 273.0 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.5 F 0.0
5:00 211.8 F 0.0
10:00 218.5 F 0.0
15:00 211.2 F 0.0
20:00 210.3 F 0.0
25:00 207.6 F 0.0
30:00 208.1 F 0.0
Drying Complete

FILLING: 1:47
HEATING: 19:49
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:58:53
CYCLE COMPLETE

09/15/2023 06:13 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____