
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 006857
10:33 29/01/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:16
PRESSURIZING 3:16
STERILIZING 4:16
136.5 C 222kPa 4:16
136.4 C 223kPa 4:46
136.4 C 223kPa 5:16
136.2 C 222kPa 5:46
136.4 C 225kPa 6:16
136.4 C 222kPa 6:46
136.3 C 223kPa 7:16
136.4 C 224kPa 7:46
136.4 C 224kPa 7:46
VENTING 7:46
AIR DRYING 8:41
CYCLE COMPLETE 8:47

Digital Signature #
EFB4B414EE360CC5
