
Midmark M11 - v1.0.0
Total Cycles: 880

Name: RITTER 1
Sterilizer ID: V2178198
06/10/2021 08:37 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 74.3 F 0.0
2:00 85.6 F 0.0
4:00 80.5 F 0.0
6:00 108.2 F 0.0
8:00 142.9 F 0.1
10:00 172.3 F 0.2
12:00 189.8 F 1.2
14:00 221.2 F 4.0
16:00 227.4 F 6.7
18:00 235.3 F 9.9
20:00 242.2 F 13.0
22:00 235.1 F 9.6
24:00 239.6 F 10.4
26:00 244.9 F 12.4
28:00 241.9 F 10.9
30:00 253.7 F 16.8
32:00 263.9 F 22.8
34:00 272.2 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 27.8
1:00 272.6 F 28.5
2:00 271.3 F 27.8
3:00 272.6 F 28.8
4:00 271.9 F 28.1

Min 271.0 F 27.7
Max 273.0 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 166.3 F 0.0
5:00 234.1 F 0.0
10:00 230.1 F 0.0
15:00 200.4 F 0.0
20:00 207.3 F 0.0
25:00 194.8 F 0.0
30:00 193.2 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 34:14
STERILIZING: 4:00
VENTING: 3:27
DRYING: 30:00
TOTAL CYCLE: 01:12:35
CYCLE COMPLETE

06/10/2021 09:49 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____