
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 001998
9:10 10/09/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:12
PRESSURIZING 3:15
STERILIZING 4:21
136.4 C 224kPa 4:21
136.2 C 223kPa 4:51
136.4 C 225kPa 5:21
136.2 C 222kPa 5:51
136.3 C 225kPa 6:21
136.1 C 224kPa 6:51
136.4 C 223kPa 7:21
136.4 C 225kPa 7:51
136.4 C 225kPa 7:51
VENTING 7:51
AIR DRYING 8:44
CYCLE COMPLETE 15:33

Digital Signature #
D5554AF7DF5EB598
