
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 007324
6:17 07/05/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:46
PRESSURIZING 3:46
STERILIZING 4:43
136.4 C 222kPa 4:43
136.1 C 221kPa 5:13
136.2 C 219kPa 5:43
136.3 C 220kPa 6:13
136.4 C 222kPa 6:43
136.3 C 221kPa 7:13
136.0 C 219kPa 7:43
136.4 C 222kPa 8:13
136.4 C 222kPa 8:13
VENTING 8:14
AIR DRYING 9:08
CYCLE COMPLETE 39:08

Digital Signature #
9262A57E212B38D8
