
Midmark M11 - v1.0.0
Total Cycles: 772

Name: RITTER 1
Sterilizer ID: V2178198
04/06/2021 02:15 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 185.5 F 0.0
2:00 180.3 F 3.0
4:00 179.2 F 7.3
6:00 208.9 F 12.5
8:00 220.1 F 7.4
10:00 235.1 F 13.0
12:00 240.0 F 10.3
14:00 234.6 F 8.1
16:00 250.7 F 15.4
18:00 263.1 F 22.5
20:00 271.3 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.5
1:00 271.7 F 28.7
2:00 271.8 F 28.9
3:00 271.9 F 29.0
4:00 271.7 F 28.9

Min 271.5 F 28.4
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 174.5 F 0.0
5:00 229.9 F 0.0
10:00 238.7 F 0.0
Dry Cycle Aborted

FILLING: 1:15
HEATING: 20:42
STERILIZING: 4:00
VENTING: 2:59
DRYING: 11:21
TOTAL CYCLE: 00:40:17

04/06/2021 02:55 PM
ACCEPTED: X
REJECTED:
Operator: U01