
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002363
9:37 05/10/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:37
PRESSURIZING 3:37
STERILIZING 4:45
136.0 C 220kPa 4:45
136.2 C 223kPa 5:15
136.4 C 224kPa 5:45
136.4 C 224kPa 6:15
136.3 C 220kPa 6:45
136.4 C 222kPa 7:15
136.4 C 223kPa 7:45
136.4 C 224kPa 8:15
136.4 C 224kPa 8:15
VENTING 8:15
AIR DRYING 9:11
CYCLE COMPLETE 9:14

Digital Signature #
4CAEF4B010963D79
