
Midmark M11 - v1.0.5
Total Cycles: 137

Name: RITTER2
Sterilizer ID: V2313228
06/24/2021 10:08 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 152.5 F 0.0
2:00 149.0 F 1.7
4:00 158.1 F 5.1
6:00 181.9 F 9.8
8:00 209.3 F 5.1
10:00 217.2 F 10.0
12:00 228.2 F 7.8
14:00 238.0 F 13.1
16:00 239.0 F 11.8
18:00 253.5 F 18.2
20:00 263.8 F 25.0
22:00 272.2 F 30.4

STERILIZING
mm:ss Degrees PSI
0:00 272.2 F 30.4
1:00 272.6 F 30.3
2:00 272.8 F 30.5
3:00 272.8 F 30.9
4:00 272.7 F 30.8

Min 272.2 F 30.2
Max 273.3 F 31.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 176.2 F 0.0
5:00 237.6 F 0.0
10:00 219.1 F 0.0
15:00 219.8 F 0.0
20:00 199.9 F 0.0
25:00 191.9 F 0.0
Dry Cycle Aborted

FILLING: 1:13
HEATING: 22:01
STERILIZING: 4:00
VENTING: 2:56
DRYING: 28:08
TOTAL CYCLE: 00:58:17

06/24/2021 11:06 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____