
Midmark M11 - v1.0.0
Total Cycles: 756

Name: RITTER 1
Sterilizer ID: V2178198
03/29/2021 12:37 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 150.8 F 0.0
2:00 155.0 F 1.7
4:00 164.3 F 5.3
6:00 179.7 F 10.1
8:00 214.5 F 6.6
10:00 221.6 F 10.9
12:00 232.8 F 7.3
14:00 243.0 F 12.4
16:00 243.8 F 11.8
18:00 256.3 F 18.7
20:00 268.0 F 25.9

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.7
1:00 272.1 F 28.9
2:00 273.1 F 29.3
3:00 271.7 F 29.0
4:00 272.5 F 29.0

Min 271.2 F 28.6
Max 273.2 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.4 F 0.0
5:00 232.6 F 0.0
10:00 249.8 F 0.0
15:00 233.5 F 0.0
20:00 193.8 F 0.0
25:00 178.6 F 0.0
30:00 168.9 F 0.0
Drying Complete

FILLING: 1:07
HEATING: 21:17
STERILIZING: 4:00
VENTING: 2:58
DRYING: 30:00
TOTAL CYCLE: 00:59:22
CYCLE COMPLETE

03/29/2021 01:36 PM
ACCEPTED: X
REJECTED:
Operator: U01