
Midmark M11 - v1.0.0
Total Cycles: 4349

Name: RITTER 1
Sterilizer ID: V2178198
02/29/2024 04:41 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 157.7 F 0.0
2:00 156.0 F 3.2
4:00 178.3 F 7.0
6:00 199.7 F 11.7
8:00 220.0 F 5.3
10:00 229.8 F 10.2
12:00 232.8 F 7.6
14:00 243.2 F 12.1
16:00 240.3 F 10.3
18:00 253.7 F 16.9
20:00 265.1 F 23.8

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.0
1:00 272.2 F 28.4
2:00 271.6 F 28.4
3:00 271.9 F 28.3
4:00 271.4 F 28.3

Min 271.1 F 27.9
Max 272.6 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 179.1 F 0.0
5:00 223.5 F 0.0
10:00 201.8 F 0.0
15:00 228.5 F 0.0
20:00 226.4 F 0.0
25:00 227.9 F 0.0
30:00 232.5 F 0.0
Drying Complete

FILLING: 1:14
HEATING: 21:47
STERILIZING: 4:00
VENTING: 3:23
DRYING: 30:00
TOTAL CYCLE: 01:00:24
CYCLE COMPLETE

02/29/2024 05:41 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____