

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R750  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000026  
9:54 09/01/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:08  
PRESSURIZING 3:13  
STERILIZING 4:11  
136.0 C 220kPa 4:11  
136.4 C 224kPa 4:41  
136.3 C 222kPa 5:11  
136.0 C 221kPa 5:41  
136.3 C 223kPa 6:11  
136.2 C 223kPa 6:41  
136.3 C 224kPa 7:11  
136.4 C 224kPa 7:41  
136.4 C 224kPa 7:41  
VENTING 7:41  
AIR DRYING 8:33  
CYCLE COMPLETE 8:42

-----  
Digital Signature #  
070E8BE34834ED0E  
-----