
Midmark M11 - v1.0.0
Total Cycles: 1731

Name: RITTER 1
Sterilizer ID: V2178198
03/21/2022 12:23 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 187.5 F 0.0
2:00 185.9 F 4.0
4:00 202.8 F 8.4
6:00 220.1 F 13.2
8:00 228.1 F 7.7
10:00 239.7 F 13.2
12:00 240.9 F 11.2
14:00 239.7 F 10.2
16:00 253.1 F 16.9
18:00 264.4 F 23.6

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.3
1:00 272.1 F 28.5
2:00 272.7 F 29.0
3:00 271.6 F 28.5
4:00 272.2 F 28.6

Min 271.5 F 28.2
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.9 F 0.0
5:00 232.6 F 0.0
10:00 225.7 F 0.0
15:00 222.5 F 0.0
20:00 219.5 F 0.0
25:00 220.8 F 0.0
30:00 222.8 F 0.0
Drying Complete

FILLING: 2:10
HEATING: 19:55
STERILIZING: 4:00
VENTING: 3:26
DRYING: 30:00
TOTAL CYCLE: 00:59:31
CYCLE COMPLETE

03/21/2022 01:22 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____