

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1691

Name: RITTER 1  
Sterilizer ID: V2178198  
03/02/2022 10:44 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 148.1 F 0.0  
2:00 149.7 F 0.3  
4:00 182.4 F 1.9  
6:00 197.8 F 6.0  
8:00 218.4 F 10.8  
10:00 222.4 F 4.9  
12:00 236.3 F 10.2  
14:00 233.7 F 8.0  
16:00 245.6 F 13.1  
18:00 243.5 F 11.8  
20:00 256.5 F 18.5  
22:00 267.8 F 25.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.0  
1:00 272.4 F 28.9  
2:00 272.1 F 28.4  
3:00 271.2 F 27.9  
4:00 271.9 F 28.7

Min 271.2 F 27.9  
Max 272.5 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 180.7 F 0.0  
5:00 222.9 F 0.0  
10:00 236.3 F 0.0  
15:00 232.1 F 0.0  
20:00 225.6 F 0.0  
25:00 225.9 F 0.0  
30:00 228.1 F 0.0  
Drying Complete

FILLING: 0:59  
HEATING: 23:22  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 01:01:29  
CYCLE COMPLETE

03/02/2022 11:45 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_