

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1967

Name: RITTER 1  
Sterilizer ID: V2178198  
05/27/2022 02:08 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 196.5 F 0.0  
2:00 186.9 F 4.3  
4:00 211.9 F 9.1  
6:00 225.2 F 8.0  
8:00 232.4 F 9.7  
10:00 235.6 F 8.6  
12:00 243.7 F 12.9  
14:00 242.4 F 11.6  
16:00 255.3 F 18.2  
18:00 266.7 F 25.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.4  
1:00 272.8 F 29.2  
2:00 271.7 F 28.4  
3:00 272.0 F 28.9  
4:00 272.4 F 29.0  
  
Min 271.2 F 28.3  
Max 272.8 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.7 F 0.0  
5:00 231.8 F 0.0  
10:00 223.3 F 0.0  
15:00 226.5 F 0.0  
20:00 229.5 F 0.0  
25:00 231.6 F 0.0  
30:00 233.4 F 0.0  
  
Drying Complete

FILLING: 2:30  
HEATING: 19:34  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 00:59:18  
CYCLE COMPLETE

05/27/2022 03:07 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_