
Midmark M11 - v1.0.0
Total Cycles: 4650

Name: RITTER 1
Sterilizer ID: V2178198
06/17/2024 04:34 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 87.2 F 0.0
2:00 87.7 F 0.2
4:00 104.6 F 0.5
6:00 129.6 F 2.2
8:00 159.4 F 4.8
10:00 187.2 F 7.6
12:00 210.2 F 11.4
14:00 212.6 F 4.1
16:00 220.3 F 7.5
18:00 230.3 F 11.4
20:00 228.3 F 6.7
22:00 239.6 F 12.1
24:00 237.7 F 9.6
26:00 249.8 F 15.2
28:00 260.0 F 20.9
30:00 269.7 F 27.1

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.8
1:00 272.6 F 29.2
2:00 271.8 F 29.1
3:00 272.4 F 29.1
4:00 271.9 F 29.2

Min 271.5 F 28.7
Max 272.8 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.7 F 0.0
5:00 216.7 F 0.0
10:00 210.9 F 0.0
15:00 214.4 F 0.0
20:00 223.7 F 0.0
25:00 231.0 F 0.0
30:00 230.8 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 31:46
STERILIZING: 4:00
VENTING: 3:50
DRYING: 30:00
TOTAL CYCLE: 01:10:29
CYCLE COMPLETE

06/17/2024 05:44 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____