
Midmark M11 - v1.0.5
Total Cycles: 1463

Name: RITTER2
Sterilizer ID: V2313228
09/07/2022 12:41 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 162.5 F 0.0
2:00 163.6 F 3.4
4:00 175.3 F 7.4
6:00 202.9 F 12.6
8:00 217.4 F 6.9
10:00 231.6 F 12.2
12:00 234.0 F 9.5
14:00 237.5 F 9.5
16:00 247.1 F 14.5
18:00 259.2 F 21.1
20:00 270.0 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 29.2
1:00 271.9 F 29.7
2:00 272.5 F 29.7
3:00 271.4 F 29.1
4:00 272.4 F 30.0

Min 271.3 F 29.1
Max 272.7 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.8 F 0.0
5:00 207.7 F 0.0
10:00 220.7 F 0.0
15:00 223.8 F 0.0
20:00 222.2 F 0.0
25:00 221.0 F 0.0
30:00 217.3 F 0.0
Drying Complete

FILLING: 1:10
HEATING: 20:47
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:59:06
CYCLE COMPLETE

09/07/2022 01:40 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____