
Midmark M11 - v1.0.5
Total Cycles: 4367

Name: RITTER2
Sterilizer ID: V2313228
02/20/2025 06:49 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 170.9 F 0.0
2:00 174.8 F 3.4
4:00 204.0 F 8.8
6:00 221.4 F 8.6
8:00 231.3 F 10.4
10:00 232.8 F 7.7
12:00 241.2 F 10.9
14:00 246.7 F 14.1
16:00 259.5 F 21.3
18:00 271.3 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.0
1:00 272.4 F 29.6
2:00 271.4 F 29.3
3:00 272.3 F 29.6
4:00 271.3 F 29.0

Min 271.3 F 28.9
Max 272.6 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.6 F 0.0
5:00 217.0 F 0.0
10:00 205.0 F 0.0
15:00 200.8 F 0.0
20:00 195.9 F 0.0
25:00 194.7 F 0.0
30:00 191.8 F 0.0
Drying Complete

FILLING: 2:22
HEATING: 18:39
STERILIZING: 4:00
VENTING: 3:31
DRYING: 30:00
TOTAL CYCLE: 00:58:32
CYCLE COMPLETE

02/20/2025 07:47 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____