
Midmark M11 - v1.0.5
Total Cycles: 3472

Name: RITTER2
Sterilizer ID: V2313228
03/22/2024 01:38 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 118.1 F 0.0
2:00 120.6 F 0.5
4:00 139.0 F 2.6
6:00 165.8 F 6.0
8:00 188.5 F 10.1
10:00 205.7 F 6.7
12:00 217.1 F 7.6
14:00 228.8 F 12.1
16:00 233.4 F 8.9
18:00 240.0 F 10.9
20:00 245.7 F 13.4
22:00 257.7 F 19.8
24:00 268.0 F 26.3

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.9
1:00 272.2 F 29.2
2:00 271.5 F 29.3
3:00 272.0 F 29.1
4:00 271.5 F 29.3

Min 271.1 F 28.7
Max 272.5 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.6 F 0.0
5:00 209.3 F 0.0
10:00 203.8 F 0.0
15:00 193.8 F 0.0
20:00 190.8 F 0.0
25:00 190.2 F 0.0
30:00 189.5 F 0.0
Drying Complete

FILLING: 0:60
HEATING: 25:20
STERILIZING: 4:00
VENTING: 3:36
DRYING: 30:00
TOTAL CYCLE: 01:03:55
CYCLE COMPLETE

03/22/2024 02:41 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____