

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4413

Name: RITTER 1  
Sterilizer ID: V2178198  
03/27/2024 03:19 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 151.1 F 0.0  
2:00 151.8 F 2.5  
4:00 157.7 F 5.5  
6:00 192.9 F 9.5  
8:00 212.2 F 10.6  
10:00 216.7 F 6.4  
12:00 228.1 F 10.6  
14:00 230.7 F 7.6  
16:00 239.8 F 11.7  
18:00 237.9 F 9.6  
20:00 249.9 F 15.4  
22:00 260.3 F 21.2  
24:00 269.9 F 27.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.6  
1:00 272.5 F 29.3  
2:00 271.5 F 28.5  
3:00 272.3 F 29.4  
4:00 272.0 F 28.9

Min 271.2 F 28.5  
Max 272.7 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 212.9 F 0.0  
5:00 217.1 F 0.0  
10:00 222.9 F 0.0  
15:00 223.5 F 0.0  
20:00 218.2 F 0.0  
25:00 216.1 F 0.0  
30:00 214.0 F 0.0  
Drying Complete

FILLING: 1:13  
HEATING: 25:03  
STERILIZING: 4:00  
VENTING: 3:45  
DRYING: 30:00  
TOTAL CYCLE: 01:04:01  
CYCLE COMPLETE

03/27/2024 04:22 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_