
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000967
9:26 09/03/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:43
PRESSURIZING 2:50
STERILIZING 4:02
136.4 C 233kPa 4:02
136.2 C 232kPa 4:32
136.1 C 229kPa 5:02
136.0 C 230kPa 5:32
136.4 C 231kPa 6:02
136.3 C 233kPa 6:32
136.4 C 231kPa 7:02
136.3 C 232kPa 7:32
136.3 C 232kPa 7:32
VENTING 7:32
AIR DRYING 8:22
CYCLE COMPLETE 8:27

Digital Signature #
6D901FCEFC7CF1DE
