

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4167

Name: RITTER 1  
Sterilizer ID: V2178198  
12/28/2023 05:17 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 169.2 F 0.0  
2:00 168.0 F 3.3  
4:00 186.6 F 7.4  
6:00 222.8 F 12.9  
8:00 228.2 F 7.4  
10:00 240.1 F 12.5  
12:00 237.2 F 9.6  
14:00 237.7 F 9.0  
16:00 248.5 F 14.5  
18:00 260.6 F 21.2  
20:00 271.7 F 28.3

STERILIZING  
mm:ss Degrees PSI  
0:00 272.4 F 28.6  
1:00 272.2 F 28.6  
2:00 272.5 F 28.7  
3:00 272.8 F 28.9  
4:00 272.5 F 29.1  
  
Min 271.5 F 28.4  
Max 272.8 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 198.0 F 0.0  
5:00 230.2 F 0.0  
10:00 227.4 F 0.0  
15:00 229.8 F 0.0  
20:00 228.5 F 0.0  
25:00 227.0 F 0.0  
30:00 214.1 F 0.0  
Drying Complete

FILLING: 1:39  
HEATING: 21:04  
STERILIZING: 4:00  
VENTING: 3:26  
DRYING: 30:00  
TOTAL CYCLE: 01:00:08  
CYCLE COMPLETE

12/28/2023 06:17 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_